



REQUEST FOR QUOTATION
PR NO. 26-09-1249

October 06, 2026

The Valenzuela Medical Center (VMC) through its Bids and Awards Committee (BAC) intends to procure the “SUPPLY, DELIVERY, ACCEPTANCE, TESTING AND COMMISSIONING OF BIOMEDICAL REFRIGERATOR (320L CAPACITY)” for the OPERATING ROOM in accordance with Section 34.1 (Small Value Procurement) of the Implementing Rules and regulations of the Republic Act No. 12009.

Please send your **price offer** for the items listed below duly signed by you or your authorized representative not later than October 09, 2026 10:00am. The BAC reserves the right to accept, reject & waive defects in the canvass.

ITEM NO.	QTY	UOM	ITEM DESCRIPTION	UNIT COST	TOTAL COST
1.	1	unit	Supply, delivery, acceptance, testing and commissioning of BIOMEDICAL REFRIGERATOR 320L Capacity	175,000.00	175,000.00
			Biomedical Refrigerator designed for safe storage of blood products, reagents, vaccines, sterilants, and sterile solutions under controlled temperature conditions. Suitable for Operating Room and Delivery Room applications requiring immediate access to temperature-sensitive materials.		
			TECHNICAL SPECIFICATION		
			External Dimensions (W × D × H): Approx. 600 × 680 × 1900 mm		
			Capacity: 320 liters		
			Temperature Range: 2°C to 8°C		
			Temperature Accuracy: ±0.1°C		
			Control System: Microprocessor-based control with large LED/LCD display		
			Alarm System: Audible and visual alarms for high/low temperature, sensor failure, and power interruption		
			Refrigeration Type: Forced air circulation system, CFC-free refrigerant (R134a or equivalent)		
			Defrost System: Automatic frost-free design		
			STRUCTURE		
			Cabinet Design: Unibody construction with rigid polyurethane insulation		
			Internal Material: High-impact polystyrene (HIPS), abrasion and acid resistant		
			External Material: Cold-rolled steel coated with anti-bacterial powder finish		
			Shelving: Four (4) adjustable anti-bacterial shelves with load capacity ≥ 30 kg each		
			Lighting: LED interior lamp for clear visibility		
			Accessories: USB port, probe access port, door lock with keys, caster wheels with brakes		
			CFC-free and environmentally safe design		
			Equipped with temperature stability and uniformity control system		
			ELECTRICAL REQUIREMENTS		

“PHIC Accredited Healthcare Provider”
“Valenzuela Medical Center... Where your health matters most”



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**



			Equipment shall compatible with hospital power supply of 220 to 240VAC, single phase, 60Hz with grounding		
			GENERAL REQUIREMENTS:		
			<i>Certification that needs to be presented during Post-Qualification/Evaluation:</i>		
			With valid FDA Certificate of Medical Device Notification (CMDN)/Certificate of Medical Device Registration (CMDR) or pending renewal at least one (1) month before expiration		
			The equipment or devices must conform to the IEC 60601 Standards and any of the following Standards:		
			a. USFDA (Food and Drug administration,US Standards b. CE(European conformity) Standards c. ISO Standards (9000, 9001, 9002) d. UL (Underwriters Laboratories) Standards e. The proposed unit shall also conform to all relevant International, National and local standards and requirements of medical device regulatory agencies. Supplier shall list in the various international standards met by the supplier		
			Manufacturer's Certificate Brand must be in the local market for at least 5 years		
			Certification that there is established Service Center in Metro Manila		
			Certificate of Comprehensive Warranty which states the following clause:		
			a. The period of warranty shall be deemed to be fully comprehensive b. All inclusive of warranty, labor, spare parts, accessories, service consumables, manufacturer's Preventive Maintenance		
			Certificates to be submitted upon delivery:		
			Warranty and Preventive Maintenance Services:		
			One (1) year comprehensive warranty on parts and services for equipment		
			Certificate of availability of spare parts within ten (10) years		
			Certification that the 95% uptime of the product is within the warranty period and that any accumulated downtime in excess of 5% shall be added to the warranty period.		
			Submit Preventive Maintenance and Calibration Schedule		
			Semi-Annual Preventive Maintenance Service and Annual Calibration during warranty period		
			Calibration Report from the Manufacturer or Verification Report from the Supplier		
			Certification that the bidder will provide a Service Unit that the end-user can use in case the equipment or any system component will be pulled-out for repair or maintenance within the warranty period.		
			Procedures to be performed upon delivery:		
			Training of End-User (Operation) and Biomedical Unit (Operation, Recommended Preventive Maintenance & Troubleshooting,)		

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**



			with Certificate of Training should be provided and should contain the following details:		
			a. Name of Trainee b. Modality, Brand, Model of Equipment c. Type of Training Conducted d. Inclusive Dates of the Training e. Name of Trainer, Date and Venue		
			The unit to be delivered must show proof that its manufacturing date is not later than CY 2024 onwards		
			Perform Electrical Safety Verification (IEC 62353)		
			Provide two (2) sets of colored Technical Manual, Hardcopy and Softcopy (Flashdrive) in English Manual for all equipment including peripherals, etc.		
			a. User's Operational Manual b. Quality and Maintenance Manual c. Service and Technical Manual		
			Other Terms & Condition:		
			Preferably with green eco-products specification (made of recyclable and non-toxic materials, energy efficient, sustainability, etc.)		
			Delivery: Sixty (60) calendar days from the date of receipt of Purchase Order (PO)		
			TOTAL		175,000.00

Copies of the following documentary requirements such as **Certificate of Registration** from BIR, **SEC/DTI**, **Business/Mayor's Permit with Official Receipts**, **Platinum PhilGEPS Certificate**, **Brochure** (if applicable), **Tax Clearance**, and **Notarized Omnibus Sworn Statement** are also required to be submitted along with your quotation/proposal.

Proposal may be submitted manually to the BAC Secretariat, Procurement Office, Valenzuela Medical Center, Padrigal Street, Karuhatan, Valenzuela City or through email at vmc_bac2026@yahoo.com. For any clarification, you may call Ms. Diana C. Pulido at telephone no. 8294-4625 or via email at vmc_bac2026@yahoo.com.


LIGAYA E. UBALDE
Head, BAC Secretariat / Procurement

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **<https://vmc.doh.gov.ph/>**