



BID BULLETIN NO. 2
13 JULY 2026

Public Bidding No. VMC-2026-059

SUPPLY, DELIVERY, ACCEPTANCE TESTING, AND COMMISSIONING OF TWO (2) UNITS DIGITAL MOBILE RADIOGRAPHY SYSTEM COMPLETELY INTEGRATED CONSOLE EACH WITH ONE (1) PIECE WIRELESS 35 X 43 CM GLASS-FREE DETECTOR FOR VALENZUELA MEDICAL CENTER (REBID)

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on July 7, 2026, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in Bidding Documents/Bid Bulletin No. 1	Amendment
SUPPLY, DELIVERY, ACCEPTANCE TESTING, AND COMMISSIONING OF TWO (2) UNITS DIGITAL MOBILE RADIOGRAPHY SYSTEM COMPLETELY INTEGRATED CONSOLE EACH WITH ONE (1) PIECE WIRELESS 35 X 43 CM GLASS-FREE DETECTOR FOR VALENZUELA MEDICAL CENTER (REBID)	
PERFORMANCE FEATURES AND SPECIFICATIONS:	PERFORMANCE FEATURES AND SPECIFICATIONS:
> Self-contained battery driven System.	> Self-contained battery driven System. Should have alarm system that can notify x-ray machine unit if the detector is left behind and when x-ray is unplugged when not in use for at least 30 minutes.
Modality Worklist > Must have softwares incorporated to the unit, such as: (Compatible to Brand being offered) a. Tube and Line Visualization Software b. Pneumothorax Visualization Software c. Bone Suppression Software d. Seam-less stitching Software (for whole spine and long bone studies) e. virtual grid software for anti-scatter correction	Modality Worklist > Must have softwares incorporated to the unit, such as: (Compatible to Brand being offered) a. Tube and Line Visualization Software. Bidder should provide published white paper or research paper for each software discussing the feature, which must be freely accessible online. b. Pneumothorax Visualization Software. Bidder should provide published white paper or research paper for each software discussing the feature, which must be freely accessible online. c. Bone Suppression Software. Bidder should provide published white paper or research paper for each software discussing the feature, which must be freely accessible online. d. Seam-less stitching Software (for whole spine and long bone studies). Bidder should provide published white paper or research paper for each software discussing the feature, which must be freely accessible online. e. virtual grid software for anti-scatter correction. Bidder should provide published white paper or research paper for each software discussing the feature, which must be freely accessible online. f. Smart Noise Cancellation Software. Bidder should provide published white paper or research paper for each software discussing the feature, which must be freely accessible online. g. Enhanced Visualization Processing Software.

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**




	Bidder should provide published white paper or research paper for each software discussing the feature, which must be freely accessible online.
MOBILE X-RAY DIMENSION:	MOBILE X-RAY DIMENSION:
> Drive Handle Height - 920 to 960 mm (Measured from floor to Top Handle)	> Drive Handle Height - 920 to 1020 mm (Measured from floor to Top Handle)
X-RAY GENERATOR:	X-RAY GENERATOR:
> mAs range: capable of 0.1 to 350	> mAs range: capable of 0.1 to 320
X- RAY TUBE MOVEMENTS:	X- RAY TUBE MOVEMENTS:
> Horizontal focal spot position (From floor): Minimum 630 - 850 mm Maximum 1340 - 1355 mm	> Horizontal focal spot position (From floor): Minimum 630 - 850 mm Maximum 1200 - 1355 mm
IMAGING STATION:	IMAGING STATION:
> Acquire and process digital images	Acquire and process digital images and provision of separate stand-alone image storage with Mini-PACS capable of storing at least 10,000 images.
ACCESSORIES:	ACCESSORIES:
b.) one (1) unit 1-shot QA phantom (for x-ray or more modality) for 1 digital portable unit - used for QC/QA	b.) one (1) unit 1-shot QA phantom (for x-ray or more modality) for 1 digital portable unit - used for QC/QA (QA Phantom for each of the machine)
GENERAL REQUIREMENTS Certification that needs to be presented during Post-Qualification:	GENERAL REQUIREMENTS Certification that needs to be presented during Post-Qualification:
NONE	The Bidder of the Medical Equipment shall comply with ISO 9001:2015 (distribution of medical devices and equipment)
NONE	The bidder/supplier must provide a certification that the digital mobile x-ray, glass-free detectors, and the software being offered are manufactured by the same manufacturer.


BID OPENING: JULY 21, 2026 @ 9:00AM

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**

