

BID BULLETIN NO. 2
06 OCTOBER 2026

Public Bidding No. VMC-2026-071
SUPPLY, DELIVERY, ACCEPTANCE TESTING, AND COMMISSIONING OF ONE (1) LOT
HYSTEROSCOPIC TOWER WITH DIAGNOSTIC, OPERATIVE AND MECHANICAL
TISSUE REMOVAL SET

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on September 29, 2026, likewise, respond to bidder’s written queries received within the prescriptive period of filing.

Posted in Bidding Documents/Bid Bulletin No. 1	Amendment
MONITOR 27" FULL HD Medical Grade Monitor -Screen resolution 1920 x 1080 -Image format 16:9 -Video inputs: 2x DVI, 1x 3G-SDI, 1x VGA, 1x S-Video, 1x Composite or its equivalent -Video outputs: 1x DVI, 1x 3G-SDI, 1x Composite Inclusion: -1 External 24 VDC Power Supply -1 Mains Cord -1 cable cover (with mounting screws)	MONITOR At least 27" FULL HD Medical Grade Monitor -Screen resolution 1920 x 1080 -Image format at least 16:9 or higher -Video inputs: 2x DVI, 1x 3G-SDI, 1x VGA, 1x S-Video, 1x Composite or its equivalent -Video outputs: 1x DVI, 1x 3G-SDI, 1x Composite Inclusion: -1 External 24 VDC Power Supply -1 Mains Cord -1 cable cover (with mounting screws)
LIGHTS SOURCE LED Light Source, 150 - 175 watts -Lamp service life: 30,000 hrs -Color Temperature: 6400K -Light Intensity adjustment: via camera head, keyboard manual via the unit -With communication bus with camera system -Class 1 Device, CF degree of safety -with 1 additional LED Light Source, 150 - 175 watts (Optional)	LIGHTS SOURCE LED Light Source, 150 - 175 watts -Lamp service life: at least 30,000 hrs -Color Temperature: 5500 - 6500K -Light Intensity adjustment: via camera head, keyboard or manual via the unit or its equivalent -With communication bus with camera system -Class 1 Device, CF degree of safety -with 1 additional LED Light Source, 150 - 175 watts (Optional)
FLUID MANAGEMENT SYSTEM 1 Unit Fluid Management System (1 irrigation set and 1 suction set) -Interdisciplinary single roller pump for irrigation or suction with glass front and touch screen -Modular design can support basic functionalities across various fields of application. -Pump can be equipped with various software modules multiple disciplines such as Lap, Hysteroscopy, Urology, Arthro and Spine. -Tubing set recognition, color coded tubing sets for clear identification -One hand installation of cassette	FLUID MANAGEMENT SYSTEM 1 Unit Fluid Management System (1 irrigation set and 1 suction set) -Interdisciplinary single roller pump for irrigation or suction with glass front and touch screen -Modular design can support basic functionalities across various fields of application. -Pump can be equipped with various software modules for multiple disciplines such as Lap, Hysteroscopy, Urology, Arthro and Spine. -Tubing set recognition, color coded tubing sets for clear identification -One hand installation of cassette or its equivalent

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



-Cartridge locking lever or its equivalent -HYS Irrigation pressure: 20 to 150 mmHg -HYS Irrigation flow rate: 200 to 600 ml/min -HYS Suction Flow: 10-180 ml/min -Height difference compensation (between the device and patient) -With irrigation tubing set, pressure controlled, reusable (1 and steam sterilizable -With irrigation tubing set, pressure controlled, sterile and single use (10 pcs) -With suction tubing set, (20 pcs)	-Cartridge locking lever or its equivalent -HYS Irrigation pressure: 20 to 150 mmHg -HYS Irrigation flow rate: 200 to 600 ml/min -HYS Suction Flow: 10-180 ml/min -Height difference compensation (between the device and patient) -With irrigation tubing set, pressure controlled, reusable (1 pc) and steam sterilizable -With irrigation tubing set, pressure controlled, sterile and single use (10 pcs) -With suction tubing set, (20 pcs)
ELECTROSURGICAL UNIT	ELECTROSURGICAL UNIT
Electrosurgical Unit for monopolar and bipolar applications in gynecology -High resolution color touch display -Wipe, impact and scratch resistant safety glass -Stored presettings, for universal use -At least 18 monopolar and bipolar cutting modes; at least 25 monopolar and bipolar coagulation modes -With specific bipolar resection mode for gynecology. -Offers real bipolar system with the current returned via the electrode and not the resectoscope sheath or its equivalent -RFID / radio frequency identification for clear identification of instrument and application with suitable cable connection -HF power output: 400W (max.) -With inclusion of two pedal footswitch	Electrosurgical Unit for monopolar and bipolar applications in gynecology -High resolution color touch display -Wipe, impact and scratch resistant safety glass -Stored presettings, for universal use -At least 18 monopolar and bipolar cutting modes; at least 25 monopolar and bipolar coagulation modes -With specific bipolar resection mode for gynecology. -Offers real bipolar system with the current returned via the electrode and not the resectoscope sheath or its equivalent -RFID / radio frequency identification for clear identification of instrument and application with suitable cable connection or its equivalent -HF power output: 400W (max.) -With inclusion of two pedal footswitch
DOCUMENTATION SYSTEM	DOCUMENTATION SYSTEM
Documentation System with 12.5" Touchscreen -Still image and video sequences capture in 2D FULL HD -Entry of patient data via keyboard or smartscreen or its equivalent -Image and video data capture via camera buttons, smart screen touch screen or keyboard -Image and video data editing module -Reference module to retrieve completed procedures Storage locations: -1TB Internal memory -USB storage media via 2.0 and 3.0 -With 1 pc USB Adaptor, for connection of recorder and connect module, for remote output	Documentation System with 12.5" Touchscreen or its equivalent -Still image and video sequences capture in 2D FULL HD -Entry of patient data via keyboard or smartscreen or its equivalent -Image and video data capture via camera buttons, smart screen touch screen or keyboard -Image and video data editing module -Reference module to retrieve completed procedures Storage locations: -1TB Internal memory -USB storage media via 2.0 and 3.0 -With 1 pc USB Adaptor, for connection of recorder and connect module, for remote output
Certification that needs to be presented during Post-Qualification	Certification that needs to be presented during Post-Qualification
Manufacturer's Certificate that the brand has been established in the local market for a minimum of ten (10) years.	Manufacturer's Certificate that the brand has been established in the local market for a minimum of five (5) years .

"PHIC Accredited Healthcare Provider"
 "Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
 Telephone Nos: **8294-6711 to 17**
 Director's Office Direct Line: **8291-4259**
 Email Address: **valgen_hosp@yahoo.com**
 Website: **https://vmc.doh.gov.ph/**




At least five (5) documentary proof of supply to hospitals or public healthcare institutions shall be submitted, such as a purchase order, sales invoice, proof of delivery, or other equivalent supporting document acceptable to the procuring entity.	At least one (1) documentary proof of supply to hospitals or public/ private healthcare institutions shall be submitted, such as a purchase order, sales invoice, proof of delivery, or other equivalent supporting document acceptable to the procuring entity.
Certificates to be submitted upon delivery:	Certificates to be submitted upon delivery:
Certificate of availability of spare parts within ten (10) years	Certificate of availability of spare parts within five (5) years


BID OPENING: OCTOBER 13, 2026 @ 9:00AM

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **<https://vmc.doh.gov.ph/>**