



BID BULLETIN NO. 2
23 JUNE 2026


Public Bidding No. VMC-2026-049
SUPPLY, DELIVERY, INSTALLATION, ACCEPTANCE TESTING, AND COMMISSIONING OF ONE (1)
BRAND NEW HIGH-END MOBILE C-ARM SYSTEM FOR VALENZUELA MEDICAL CENTER


Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on June 16, 2026, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in Bidding Documents/Bid Bulletin No. 1	Amendment
General Features	General Features
Fluoroscopic and Radiographic kV range = 40 kVp to 120 kVp	Fluoroscopic and Radiographic kV range = 40 kVp to 125 kVp
Digital Cine pulse: can reach 150mA - 250mA	Digital Cine pulse: within 150mA - 250mA
Digital spot: can reach 75mA - 250mA	Digital spot: within 75mA - 250mA
C-Arm Dimensions	C-Arm Dimensions
Length: 200-215 cm	Length: 195 to 215 cm
Depth in arc: 65-70 cm	Depth in arc: 65 to 72 cm
Movement:	Movement:
a. Orbital rotation: at least 148 degrees	a. Orbital rotation: at least 146 degrees
Width: 30 to 40 inches	Width: 29 to 40 inches
Weight: 136 to 204 kg	Weight: 136 to 230 kg
Workstation Display	Workstation Display
g. Touch Control Panel: should have easy access function, including X-ray, live zoom, digital pen, measurements, annotations, laser aimer, image directory, and other functions.	g. Touch Control Panel: should have easy access function, including X-ray, live zoom, digital pen or digital graphical overlay feature, measurements, annotations, laser aimer, image directory, and other functions.
Others:	Others:
- Minimal difference spatiotemporal filter	- Minimal difference spatiotemporal filter or software for filtering and reducing noise or its equivalent
- Motion tolerant subtraction	- Motion tolerant subtraction or automatic pixel shift or its equivalent
Patient Privacy control blank screen	Patient Privacy control blank screen or Advanced Security Package to protect imaging system and patient data (HIPAA) or Radiation lock or its equivalent

BID OPENING: JUNE 30, 2026 @ 9:00AM

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:

MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:

MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: Padrigal St., Karuhatan, Valenzuela City, 1441
Telephone Nos: 8294-6711 to 17
Director's Office Direct Line: 8291-4259
Email Address: valgen_hosp@yahoo.com
Website: <https://vmc.doh.gov.ph/>