



BID BULLETIN NO. 1
29 JULY 2026

Public Bidding No. VMC-2026-062
SUPPLY, DELIVERY, ACCEPTANCE TESTING, AND COMMISSIONING OF VARIOUS LABORATORY EQUIPMENT FOR HISTOPATHOLOGY, BLOOD BANK AND MOLECULAR LABORATORY – BATCH 4

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the **Pre-bid** Conference held on July 29, 2026, likewise, respond to bidder’s written queries received within the prescriptive period for filing.

Posted in Bidding Documents		Amendment
SUPPLY, DELIVERY, ACCEPTANCE TESTING, AND COMMISSIONING OF VARIOUS LABORATORY EQUIPMENT FOR HISTOPATHOLOGY, BLOOD BANK AND MOLECULAR LABORATORY – BATCH 4		
1	Supply, Delivery Acceptance, Testing and Commissioning of Cartridge-Based Nucleic Acid Amplification Test Using Real-Time PCR Technology FOR MULTI- DISEASE TESTING OF MTB- RIF, HIV AND CO- INFECTIONS	
	I. Equipment Specification:	
	7. with Laptop, printer, and barcode scanner for cartridge identification or sample (as per manufacturers configuration)	7. with Laptop OR DESKTOP , printer, and barcode scanner for cartridge identification or sample (as per manufacturers configuration)
	8. With free 10 cartridge of SARS-CoV-2/Flu A/Flu B/RSV Plus and HPV testing with sample collection kit	8. With free 10 cartridge of SARS-CoV-2/Flu A/Flu B/RSV Plus and 10 cartridge of HPV testing with sample collection kit
2	SUPPLY, DELIVERY, INSTALLATION, ACCEPTANCE TESTING AND COMMISSIONING OF FOUR(4) BODY CAPACITY MORGUE FREEZER	
	MORGUE FREEZER	
	I. Equipment Specification:	
	1. Exterior Material : Made of stainless steel AISI-304 plate.	1. Exterior Material : Made of stainless steel AISI-304 (WITH CERTIFICATION FROM THE MANUFACTURER) plate.
	3. Interior material (Flooring, Walls and Ceiling): stainless steel finish	3. Interior material (Flooring, Walls and Ceiling): stainless steel AISI-304 (WITH CERTIFICATION FROM THE MANUFACTURER)
	8. Tray dimension : Length : at least 81.1 in. Width : at least 23 in.	8. Tray dimension : Length : at least 76.8 in. Width : at least 23 in.
	9. Four (4) telescopic rack assembly with nylon rollers that fully extends the body tray from end to end, carrying capacity - at least 100 kilograms.	9. Four (4) telescopic rack OR ROLLER TRACK assembly with nylon rollers OR SLIDE ROLLER that fully extends the body tray from end to end, carrying capacity - at least 100 kilograms.
	13. Temperature range : up to 0°C/-18°C to -25°C	13. Temperature range : up to 0°C to -25°C
	20. With built-in battery (back up system) minimum of 2 hours	20. With built-in / EXTERNAL battery FOR DISPLAY BOARD TEMPERATURE minimum of 2 hours
	21. Refrigerating Plant: Bodywork- in painted steel plate and structured with removable panels for an easy maintenance. Compressor: it's gas proof and cooled by the suctioned gas; it is equipped with an integral protecion engine with electronically adjusted themistors. Ecological gas. The gas spreading is carried out using a capillary tube or the thermostatic valve. The nominal powers are indicated in the proper illustration charts.	21. Refrigerating Plant: Bodywork- in painted steel plate / STAINLESS STEEL PLATE and structured with removable panels for an easy maintenance. Compressor: it's gas proof and cooled by the suctioned gas; it is equipped with an integral protecion engine with electronically adjusted themistors. Ecological gas. The gas spreading is carried out using a capillary tube or the

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**



		thermostatic valve. The nominal powers are indicated in the proper illustration charts.
	III. General Requirements:	
	Certification that needs to be presented during Post-Qualification	
	With valid FDA Certificate of Medical Device Notification (CMDN)/Certificate of Medical Device Registration (CMDR) or pending renewal at least one (1) month before expiration Submit the following as proof of renewal application: - Proof of payment - Acknowledgement Receipt (e.g. FDA DocTrack)	REMOVE ATTACH FDA MEMORANDUM NON-REGISTRABLE MEDICAL DEVICE WITH PROOF OF COMMUNICATION WITH FDA (IF APPLICABLE) OR CERTIFICATE FROM THE MANUFACTURER
	Bidder must undertake a site inspection. A site inspection certificate shall be issued as proof of compliance. This certificate shall be submitted with the bid. (Issued by Engineering Department of VMC)	Bidder must undertake a site inspection. A site inspection certificate shall be issued as proof of compliance. This certificate shall be submitted with the bid. (Issued by Engineering Department of VMC) (TECHNICAL REQUIREMENT)
	Other Terms & Condition:	
	Preferably with green eco-products specification (made of recyclable and non-toxic materials, energy efficient, sustainability, etc.)	Preferably with green eco-products specification (made of recyclable and non-toxic materials, energy efficient, sustainability, etc.) UPON DELIVERY


BID OPENING: AUGUST 11, 2026 @ 9:00AM

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center... Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**