



BID BULLETIN NO. 1
01 JULY 2026
Public Bidding No. VMC-2026-056
SUPPLY AND DELIVERY OF VARIOUS LABORATORY SUPPLIES AND REAGENTS
FOR CLINICAL MICROSCOPY, SEROLOGY AND HEMATOLOGY
(2ND SEMESTER OF CY 2026)

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on July 1, 2026, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Item No.	Posted in Bidding Documents	Item No.	Amendment
	CLINICAL MICROSCOPY SECTION		CLINICAL MICROSCOPY SECTION
1	Specification for No. 1 :For Fully-Automated Urine Physical, Chemical and Sediment Analyzer Reagent (free use of machine/tie-up): - Fully Automated Urine Physical and Chemistry Analyzer Specification: Should utilize Reflectance Photometry for test strip analysis and Refractometry for measuring Specific Gravity. It must achieve a minimum throughput of 200 samples per hour, with an aspiration volume of no more than 1mL of urine sample. Integration with a urine sediment analyzer is required. Additionally, consideration will be given to systems that offer extra parameters, provided they are supported by validated protocols. - Fully Automated Urine Sediment Analyzer Specification: Digital Imaging analyzer capable of processing STAT samples, with an aspiration volume of no more than 1mL of urine and a minimum throughput of 100 samples per hour. - The system should be able to identify and differentiate various formed elements in urine, including Red Blood Cells (RBC), White Blood Cells (WBC) and WBC clumps, Epithelial cells (squamous, non-squamous, transitional, and renal), casts (hyaline and pathologic), bacteria, crystals, yeasts, sperm cells, and mucus threads. - Strongly preferred capable of analyzing body fluids (with FDA certification) and provide information on RBC morphology, specifically distinguishing between isomorphic and dysmorphic RBC. - Other requirements: >Provide consumables including disposable khan tubes or its equivalent, urine	1	Specification for No. 1 :For Fully-Automated Urine Physical, Chemical and Sediment Analyzer Reagent (free use of machine/tie-up): - Fully Automated Urine Physical and Chemistry Analyzer Specification: Should utilize Reflectance Photometry for test strip analysis and Refractometry for measuring Specific Gravity. It must achieve a minimum throughput of 200 samples per hour, with an aspiration volume of 1mL-2.5ml of urine sample. Integration with a urine sediment analyzer is required. Additionally, consideration will be given to systems that offer extra parameters, provided they are supported by validated protocols. - Fully Automated Urine Sediment Analyzer Specification: Flow Cytometry or cuvette-based with Digital Imaging analyzer capable of processing STAT samples, with an aspiration volume of no more than 1mL of urine and a minimum throughput of 100 samples per hour. - The system should be able to identify and differentiate various formed elements in urine, including Red Blood Cells (RBC), White Blood Cells (WBC) and WBC clumps, Epithelial cells (squamous, non-squamous, transitional, and renal), casts (hyaline and pathologic), bacteria, crystals, yeasts, sperm cells, and mucus threads. - can run urine osmolality - has a sample premixing sequence before aspiration of sample - Strongly preferred capable of analyzing body fluids (with FDA certification) and provide information on RBC morphology, specifically distinguishing between isomorphic and dysmorphic RBC. - Other requirements: >Provide consumables

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**



Item No.	Posted in Bidding Documents	Item No.	Amendment
	chemistry and sediments controls, and distilled water (if needed). >Complete set of computer and printer with scanner (including ink). >Result printout customizable to end-user's format; >Can generate printout of summary of reports run daily as specified by end-user; >LIS ready - The winning bidder should shoulder any expenses related to LIS connectivity. LOT BID No. 1		including disposable khan tubes or its equivalent, urine chemistry and sediments controls, and distilled water (if needed). >Complete set of computer and printer with scanner (including ink). >Result printout customizable to end-user's format; >Can generate printout of summary of reports run daily as specified by end-user; >LIS ready - The winning bidder should shoulder any expenses related to LIS connectivity. LOT BID No.
	HEMATOLOGY SECTION		HEMATOLOGY SECTION
16	For Fully-automated Hematology Analyzer (6-part diff) Reagent Free use of two machine with complete set of computer and printer ; - Result printout customizable to end-user's format; - Can generate printout of summary of reports run daily as specified by end-user; - LIS ready, the winning bidder should shoulder any expenses related to LIS connectivity. - Throughput : 80-100 samples/hour in WB mode with autoloader, closed system. - Should ran stat mode. - Aspiration volume : minimum of at least 70-90ul in consideration of pedia and neonatal patients. - Should generate 6 part differential including immature cells (must be FDA certified). - Should be capable of running other body fluids (must be FDA certified) with volume aspiration of at least 70-90 ul. - Should generate auto NRBC correction. - Automated Reticulocyte count capable. - Capable of platelet flourescence. - Control's shelf life must be at least 1 month upon delivery. - Should be able to generate printable LJ chart of controls. >All information and specification must have supporting documents. - Automated registry of controls. - Principle : Combination of Fluorescence Flow Cytometry and Electrical Impedance method with Quality Control Flagging.	16	For Fully-automated Hematology Analyzer (6 to 8-part diff) Reagent Free use of two machine with complete set of computer and printer ; - Result printout customizable to end-user's format; - Can generate printout of summary of reports run daily as specified by end-user; - LIS ready, the winning bidder should shoulder any expenses related to LIS connectivity. - Throughput : 80-120 samples/hour in WB mode with autoloader, closed system. - Should ran stat mode. - Aspiration volume : minimum of at least 70-90ul (for adult) and 25ul (for pedia and neonatal) - Should generate 6 part differential including immature cells (must be FDA certified). - Should be capable of running other body fluids (must be FDA certified) with volume aspiration of at least 70-90 ul. - Should generate auto NRBC correction. - Automated Reticulocyte count capable. - Capable of platelet flourescence. - Control's shelf life must be at least 1 month upon delivery. - Should be able to generate printable LJ chart of controls. >All information and specification must have supporting documents. - Automated registry of controls. - Principle : Combination of Fluorescence Flow Cytometry and Electrical Impedance method with Quality Control Flagging.

"PHIC Accredited Healthcare Provider"

"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**



Item No.	Posted in Bidding Documents	Item No.	Amendment
	- 24/7 after service required upon machine malfunction or failure. - Set of reagents must consists of Diluent, Lysing reagents, staining reagents, controls. - At least 500 test/set excluding controls and calibration running. LOT BID item no. 16		- 24/7 after service required upon machine malfunction or failure. - Set of reagents must consists of Diluent, Lysing reagents, staining reagents, controls. - At least 500 test/set excluding controls and calibration running. LOT BID item no. 16

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


Aileen S. Cali
BAC Secretariat

Noted by:


EDSEL S. MARTIN
Chairperson, BAC I

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**