



BID BULLETIN NO. 1
7 JULY 2026

Public Bidding No. VMC-2026-059

SUPPLY, DELIVERY, ACCEPTANCE TESTING, AND COMMISSIONING OF TWO (2) UNITS DIGITAL MOBILE RADIOGRAPHY SYSTEM COMPLETELY INTEGRATED CONSOLE EACH WITH ONE (1) PIECE WIRELESS 35 X 43 CM GLASS-FREE DETECTOR FOR VALENZUELA MEDICAL CENTER (REBID)

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the Pre-bid Conference held on July 7, 2026, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in Bidding Documents	Amendment
SUPPLY, DELIVERY, ACCEPTANCE TESTING, AND COMMISSIONING OF TWO (2) UNITS DIGITAL MOBILE RADIOGRAPHY SYSTEM COMPLETELY INTEGRATED CONSOLE EACH WITH ONE (1) PIECE WIRELESS 35 X 43 CM GLASS-FREE DETECTOR FOR VALENZUELA MEDICAL CENTER (REBID)	
BID DATA SHEET ITB Clause 5.4 The Bidder's SLCC similar to the contract to be bid should have been completed within ten (10) years prior to the deadline for the submission and receipt of bids. - must be equivalent to at least fifty percent (50%) of the Total ABC of the Project.	BID DATA SHEET ITB Clause 5.4 The Bidder's SLCC similar to the contract to be bid should have been completed within ten (10) years prior to the deadline for the submission and receipt of bids. - must be equivalent to at least fifty percent (50%) of the Total ABC of the Project. - Aggregated contracts are not allowed.
PERFORMANCE FEATURES AND SPECIFICATIONS:	PERFORMANCE FEATURES AND SPECIFICATIONS:
> Time to display preview image 4 sec or less.	> Time to display preview image 5 sec or less.
MOBILE X-RAY DIMENSION:	MOBILE X-RAY DIMENSION:
> Height from Floor to Top of Tube Head Column Extended - 2100 to 2200 mm	> Height from Floor to Top of Tube Head Column Extended - up to 2200 mm
> Wheels - 2 Rear: up to 330 mm Diameter , Wheels - 2 Front: up to 150 mm Diameter - With automatic and manual brake locking system or manufacturer's standard	> Wheels - 2 Rear: up to 330 mm Diameter , Wheels - 2 Front: up to 155 mm Diameter - With automatic and manual brake locking system or manufacturer's standard
X-RAY GENERATOR:	X-RAY GENERATOR:
> mAs range: capable of 0.1 to 320	> mAs range: capable of 0.1 to 350
TYPE OF POWER SOURCE:	TYPE OF POWER SOURCE:
> Battery lead acid: high-weighted battery is needed for stability of voltage output, which is critical during high-load exposures.	> Battery lead acid/Lithium-ion: high-weighted battery is needed for stability of voltage output, which is critical during high-load exposures.
MANUFACTURING COMPANY/BRAND	MANUFACTURING COMPANY/BRAND
> Payment for the application for License and transportation expenses and per diem of the FDA-CDRRHR Medical Physics Team shall be	> Payment for the application for License and transportation expenses and per diem of the FDA-CDRRHR Medical Physics Team shall be

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**



shouldered by the bidder and every time that x-ray tube is replaced	shouldered by the bidder and every time that x-ray tube is replaced within warranty period
GENERAL REQUIREMENTS Certification that needs to be presented during Post-Qualification:	GENERAL REQUIREMENTS Certification that needs to be presented during Post-Qualification:
Notarized Certificate of Distributorship from the Principal and Distributor at least five (5) years and has a local presence sales and support	Notarized Certificate of Distributorship from the Principal at least five (5) years and has a local presence sales and support


BID OPENING: JULY 21, 2026 @ 9:00AM

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center...Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**