



BID BULLETIN NO. 2
9 JUNE 2026

Public Bidding No. VMC-2026-047

SUPPLY, DELIVERY, ACCEPTANCE TESTING, AND COMMISSIONING OF TWO (2) UNITS DIGITAL MOBILE RADIOGRAPHY SYSTEM COMPLETELY INTEGRATED CONSOLE EACH WITH ONE (1) PIECE WIRELESS 35 X 43 CM GLASS-FREE DETECTOR FOR VALENZUELA MEDICAL CENTER

Issued pursuant to Sec. 51.5 of the IRR of RA 12009 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by the prospective bidders during the **Pre-bid** Conference held on June 3, 2026, likewise, respond to bidder's written queries received within the prescriptive period for filing.

Posted in Bidding Documents	Amendment
ENVIRONMENTAL CONDITIONS: - Transport and storage temperature -20 to 60°C - Operating Temperature: 10 to 40°C	ENVIRONMENTAL CONDITIONS: - Transport and storage temperature: within -20 to 60°C - Operating Temperature: within 10 to 40°C
MOBILE X-RAY DIMENSION: > Wheels - 2 Rear: up to 330 mm Diameter , Wheels - 2 Front: up to 150 mm Diameter - With automatic and manual brake locking system	MOBILE X-RAY DIMENSION: > Wheels - 2 Rear: up to 330 mm Diameter , Wheels - 2 Front: up to 150 mm Diameter - With automatic and manual brake locking system or manufacturer's standard
X-RAY GENERATOR: > Battery powered generator, 240 Volts direct current	X-RAY GENERATOR: > Battery powered generator, 240 Volts direct current or manufacturer's standard
X- RAY TUBE MOVEMENTS: > Column rotation range: ±270 Degrees > Tube angulation: +90 to -20 Degrees	X- RAY TUBE MOVEMENTS: > Column rotation range: ±270 Degrees or manufacturer's standard > Tube angulation: +90 to -20 Degrees or manufacturer's standard
DIGITAL DETECTOR: > "Smart" battery technology prevents overcharge > Must have a sleep mode capability	DIGITAL DETECTOR: > "Smart" battery technology prevents overcharge or its equivalent > Must have a sleep mode capability or its equivalent


BID OPENING: JUNE 17, 2026 @ 9:00AM

All other provisions of the bidding documents not affected by revisions included in this Bid Bulletin shall remain in effect. Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

Prepared by:


MS. KRISTINE JOY MANUEL
BAC(Secretariat

Noted by:


MS. RUBY S. GURREA, RN, MAN
Chairperson, BAC II

"PHIC Accredited Healthcare Provider"
"Valenzuela Medical Center... Where your health matters most"



Address: **Padrigal St., Karuhatan, Valenzuela City, 1441**
Telephone Nos: **8294-6711 to 17**
Director's Office Direct Line: **8291-4259**
Email Address: **valgen_hosp@yahoo.com**
Website: **https://vmc.doh.gov.ph/**